

060AAS Series

- Wide Range AC Input
- IT and Medical Safety Approvals
- Energy Star Compliance (Level V)
- Low Earth Leakage Current (<0.1mA)
- 2 Year Warranty
- Risk Management Report Available



Specifications

Input Voltage Range	90-264Vac
Frequency	47-63Hz
Max Input Current	1.5A at 90Vac, 0.75A at 264Vac
Max Inrush Current	50A at 115Vac, 100A at 230Vac
Average Efficiency	>87%
No Load Power Consumption	<0.5W
Hold Up Time	10ms at 115Vac typical
Input Protection	Internal T3.15A / 250V fuse in line and neutral
Earth Leakage Current	<0.1mA normal condition, <0.3mA single fault condition at 264Vac/60Hz
Output Voltage	12-24VDC (See table)
Max Power	60W
Ripple & Noise	<1% (25°C with 10uF/0.47uF and 20MHz bandwidth)
Line Regulation	±0.5%
Load Regulation	±5%
Over Voltage Protection	<150%
Over Current Protection	<180%
Shock (non operation)	50G, 11ms, 3 shocks for each direction
Vibration (non operation)	5-500Hz, 2.09Grms, 20 minute for each three axis
Operating Temperature	0 to 40°C
Storage Temperature	-40 to 85°C
Operating Humidity	10 to 90% RH, non-condensing
Storage Humidity	5 to 95% RH, non-condensing
MTBF	1218 kHrs based on Telecordia SR-332 at 25°C, 230 Vac and 100% of rated load
Dimensions (WxLxH)	62.0x135.0x34.1 mm(2.4x5.3x1.3 in)
Weight (typical)	0.25kg without cable
Emissions	EN55011/22/FCC conducted / radiated class B
Harmonic Current	NA
Voltage Flicker	EN61000-3-3
ESD Immunity	EN61000-4-2, ±6kV contact / ±8kV air Perf Criteria A
Radiated Immunity	EN61000-4-3, 3V/m Perf Criteria A
EFT	EN61000-4-4, level 3 (2kV) Perf Criteria A
Surge	EN61000-4-5, DM ±1kV / CM ±2kV Perf Criteria A
Conducted Immunity	EN61000-4-6, 3Vrms Perf Criteria A
Magnetic Field	EN61000-4-8, 3A/m, Perf Criteria A
Voltage Dips	EN61000-4-11 30% 10ms Criteria A ; 60% 100ms and 100% 5000ms Criteria B
Operating Altitude	3000m under IEC60601-1:2005, 5000m under IEC60950-1 (Ed.2,2005)
Safety Compliance	IEC60601-1:2005, EN60601-1:2006, CAN/CSA-C22.2 No. 60601-1:08
	ANSI/AAMI ES60601-1:2005
	IEC60950-1 (Ed.2,2005), GB4943.1-2011, GB9254-2008, GB17625.1-2003

060AAS Series

Output Rating


Output Voltage	Output Current	Output Power	Model Name
12V	5.0A	60W	MDS-060AAS12 B ¹ A ²
15V	4.0A	60W	MDS-060AAS15 B ¹ A ²
19V	3.2A	60W	MDS-060AAS19 B ¹ A ²
24V	2.5A	60W	MDS-060AAS24 B ¹ A ²

Notes

- AC Inlet notation, B = IEC320-C14 ; D = IEC320-C6
- DC plug/cable length notation, A = 4 pin DIN type/1200mm; B = 4 pin DIN type/1800mm;
E = barrel type 5.5*2.5, L=10mm/1200mm; F = barrel type 5.5*2.5, L=10mm/1800mm;
J = barrel type 5.5*1.7, L=10mm/1200mm; K = barrel type 5.5*1.7, L=10mm/1800mm

Mechanical Drawing

Dimensions (WxLxH): 62.0x135.0x34.1 mm



Unit: mm



**Стандарт
Электрон
Связь**

Мы молодая и активно развивающаяся компания в области поставок электронных компонентов. Мы поставляем электронные компоненты отечественного и импортного производства напрямую от производителей и с крупнейших складов мира.

Благодаря сотрудничеству с мировыми поставщиками мы осуществляем комплексные и плановые поставки широчайшего спектра электронных компонентов.

Собственная эффективная логистика и склад в обеспечивает надежную поставку продукции в точно указанные сроки по всей России.

Мы осуществляем техническую поддержку нашим клиентам и предпродажную проверку качества продукции. На все поставляемые продукты мы предоставляем гарантию .

Осуществляем поставки продукции под контролем ВП МО РФ на предприятия военно-промышленного комплекса России , а также работаем в рамках 275 ФЗ с открытием отдельных счетов в уполномоченном банке. Система менеджмента качества компании соответствует требованиям ГОСТ ISO 9001.

Минимальные сроки поставки, гибкие цены, неограниченный ассортимент и индивидуальный подход к клиентам являются основой для выстраивания долгосрочного и эффективного сотрудничества с предприятиями радиоэлектронной промышленности, предприятиями ВПК и научно-исследовательскими институтами России.

С нами вы становитесь еще успешнее!

Наши контакты:

Телефон: +7 812 627 14 35

Электронная почта: sales@st-electron.ru

Адрес: 198099, Санкт-Петербург,
Промышленная ул, дом № 19, литера Н,
помещение 100-Н Офис 331